



جمهوری اسلامی ایران  
Islamic Republic of Iran



## گواهینامه تایید صلاحیت آزمایشگاه Laboratory Accreditation Certificate

The National Accreditation Center of Iran (NACI)  
herewith confirms that :

**Azmouneh Foulad Consulting  
Engineers Calibration Laboratory**

**Address:** Azmouneh Foulad Bld, Next to the Telecom  
Bld, B7 Area, Fouladshahr, Isfahan, I.R.IRAN

**Tel :** +98(31)52629627

**Fax :** +98(31) 52629628

**Web Site :** www.azmouneh .com

Has fulfilled the **ISIRI-ISO/IEC 17025**.  
And is competent to carry out  Test  Calibration  
services according to accreditation scope are listed in  
1 page/s of annex.

**NACI Registration No :** NACI/Lab/289

**Initial Accreditation Date and Place:** 2012.04.18-Tehran

**Renewal Date :** 2016.08.06

**Expiry Date :** 2018.08.14

Validity of Accreditation Certificate depends on  
continuity of compliance with the relevant requirements  
and obtaining the approval based on the annual  
surveillance assessment.

**N. Pirouzbakht**  
PRESIDENT, IRAN ACCREDITATION COUNCIL

**A.R. Khakifirooz**  
NACI PRESIDENT

مرکز ملی تایید صلاحیت ایران بدین وسیله تایید می نماید که :

**آزمایشگاه کالیبراسیون مهندسی مشاور آزمون  
فولاد**

نشانی: اصفهان، فولادشهر، محله B7، جنب مخابرات، ساختمان آزمون  
فولاد

تلفن: ۵۲۶۲۹۶۲۷ - ۳۱

دورنگار: ۵۲۶۲۹۶۲۸ - ۳۱

سایت اینترنتی: www.azmouneh .com

الزامات استاندارد ایران - ایزو/آی ای سی ۱۷۰۲۵ را رعایت نموده است.  
و صلاحیت انجام خدمات آزمون  کالیبراسیون  مطابق دامنه کاربردی  
که جزئیات آن در ۱ برگ پیوست آمده است را داراست.

شماره گواهینامه تایید صلاحیت : NACI/Lab/289

تاریخ و محل صدور اولیه گواهینامه : ۱۳۹۱/۰۱/۳۰ - تهران

تاریخ صدور مجدد گواهینامه : ۱۳۹۵/۰۵/۱۶

تاریخ خاتمه اعتبار گواهینامه : ۱۳۹۷/۰۵/۲۳

حفظ اعتبار در طول دوره منوط به استمرار انطباق با ضوابط مربوطه و اخذ  
تاییدیه در ارزیابی های مراقبتی سالانه است.

نیره پیروزبخت  
رئیس شورای تایید صلاحیت ایران

**علیرضا خاکی فیروز**  
رئیس مرکز ملی تایید صلاحیت ایران



جمهوری اسلامی ایران  
Islamic Republic of Iran



گواهینامه تایید صلاحیت آزمایشگاه  
Laboratory Accreditation Certificate

پیوست

دامنه کاربرد تایید صلاحیت آزمایشگاه کالیبراسیون مهندسین مشاور آزمون فولاد

ردیف	کمیت، دستگاه اندازه گیری، سنجه مادی	گستره	بهترین توان اندازه گیری آزمایشگاه (±)
۱	نیروی دستگاههای آزمون فشاری و کششی	Up to 100 t	0.5%
۲	نیروی دستگاههای آزمون فشاری	100 to 280 t	4.26%
۳	اندازه گیری سرعت دستگاههای آزمون کششی	Up to 200 mm/min	0.2 mm/min
۴	اندازه گیری جابجایی دستگاههای آزمون کششی	Up to 100 mm	0.2 mm

نیره پیروزبخت  
رئیس شورای تایید صلاحیت ایران

علیرضا خاکی فیروز  
رئیس مرکز ملی تایید صلاحیت ایران

نیره





جمهوری اسلامی ایران  
Islamic Republic of Iran



گواهینامه تایید صلاحیت آزمایشگاه  
Laboratory Accreditation Certificate

Annex  
Accreditation Scope of Azmouneh Foulad Consulting Engineers Calibration  
Laboratory

No.	Quantity, Measuring Instrument, Material Measure	Range	Best Measurement Capability lab( $\pm$ )
1	Tension / Compression testing machines-calibration of the force measuring system	Up to 100 t	0.5%
2	Compression testing machines-calibration of the force measuring system	100 to 280 t	4.26%
3	Tension testing machines-calibration of the speed measuring system	Up to 200 mm/min	0.2 mm/min
4	Tension testing machines-calibration of the movement measuring system	Up to 100 mm	0.2 mm

A.R. Khakifirooz  
NACI PRESIDENT

*Khakifirooz*

N. Pirouzbakht  
PRESIDENT, IRAN ACCREDITATION COUNCIL